



Product Service

# EC-Attestation of Conformity

No. N8 09 05 48128 027

**Holder of Certificate:** Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd,Binjiang District  
310053 Hangzhou  
PEOPLE'S REPUBLIC OF CHINA

**Product:** Laboratory Equipment  
Fluorescent Quantitative Detection System

**Model(s):** FQD-96A

**Parameters:**

Rated voltage:	100-240 V~
Rated frequency:	50/60Hz
Rated input power:	600W
Protection class:	I
Degree of protection:	ordinary

**Tested according to:** EN 61010-2-010:2003

This EC-Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

**Test report no.:** 704200927301-00



**Date,** 2009-06-02

**CE** After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.



Product Service

# EC-Attestation of Compliance

No. E8N 09 04 48128 025

**Holder of Certificate:** Hangzhou Bioer Technology Co., Ltd.

1192BinAn Rd, Binjiang District  
310053 Hangzhou  
PEOPLE'S REPUBLIC OF CHINA

**Name of Object:** Laboratory Equipment  
Fluorescent Quantitative Detection System

**Model(s):** FQD-96A

**Description of Object:**

Rated voltage: AC 220-240V  
Rated frequency: 50Hz  
Rated input power: 600W  
Protection class: I

**Tested according to:** EN 61326-1:2006

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

**Test report no.:** 708880927301-00

**Date,** 2009-05-08

*Linda Grün*



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.