

WARRANTY

Products are guaranteed to conform to the quality and content indicated on each vial and external labels during their shelf life. BIOTOOLS obligation and purchaser's rights under this warranty are limited to the replacement by BIOTOOLS of any product that is shown defective in fabrication, and that must be returned to BIOTOOLS, freight prepaid, or at BIOTOOLS' option, replacement of the purchasing price. Any complaint on damaged goods during transport must be directed to the handling or transport agent.

Product for Research Use Only. This product must be used by qualified professionals only. It is the responsibility of the user to ascertain that a given product is adequate for a given application. Any product not fulfilling the specifications included in the product sheet will be replaced. This warranty limits our responsibility to the replacement of the product. No other warranties, of any kind, express or implied, including, without limitation, implicit warranties of commercialisation ability or adequacy for a given purpose, are provided by BIOTOOLS. BIOTOOLS will not be held responsible for any direct, indirect, consequential or incidental damage resulting of the use, misuses, results of the use or inability to use any product.

Produced by:

BIOTOOLS, Biotechnological & Medical Laboratories, S.A. have been evaluated and certified to accomplish ISO 9001:2008 requirements for the following activities: Research and development of biotechnology products and manufacture of biotechnology and in vitro products. Valle de Tobalina – 52 – Nave 39, 28021 Madrid – Spain

© 2008 BIOTOOLS, Biotechnological & Medical Laboratories, S.A. All rights reserved



BIOTOOLS B&M Labs, S.A.
Valle de Tobalina - 52 - Nave 39
28021 Madrid
Spain

Tel. (34) 91 710 00 74
Fax (34) 91 505 31 18
E-mail: info@biotools.eu
www.biotools.eu



BIOTOOLS
BIOTOOLS B & M LABS. S.A.

CERTAMP COMPLEX

Ref.	FORMAT	CONTENT
10.342	250 rxns	CERTAMP Complex
10.343	500 rxns	CERTAMP Complex

Store at -20°C

Notice to users: Some of the applications which may be performed with this product are covered by applicable patents in certain countries. The purchase of this product does not include or provide a license to perform patented applications. Users may be required to obtain a license depending on the country and/or application.

Ed 22- February 2016

1. GENERAL INFORMATION

CERTAMP Complex has been designed for the optimisation and improvement of amplification reactions of low-copy-number targets in complex DNA backgrounds. An exclusive combination of proof-reading and non proof-reading polymerases, cofactors and an especial buffer, provides robust amplification of complex PCR.

RESEARCH USE ONLY

Genomic DNA, templates rich in GC or with secondary structure are difficult to amplify by standard PCR techniques. The main reason of the difficulty is that when conventional polymerases (without proof-reading activity) misincorporate a nucleotide, they tend to stall and either fall off the template resulting in a truncated product or eventually continue in the product sequence. As a result, these enzymes have fairly high error rates and produce relatively low yields of largely incomplete products. Secondary structure or GC-rich templates cause the polymerase to fall off the template before completely duplicating the target sequence.

Certamp Complex is able to overcome the deficiencies of conventional polymerases largely through the combination of proof-reading and non proof-reading enzymes. The proof-reading polymerase provides 3' to 5' exonuclease activity that removes misincorporated bases, thus allowing the conventional polymerase, which has higher synthesis rates than the proof-reading ones, to continue copying the target DNA. The presence of both enzymes significantly improves fidelity and processivity, which results in high yields of highly accurate, full-length products. Polymerase mixes allow highly sensitive and accurate amplification of virtually any target, regardless of size or complexity.

This robust enzyme system also minimizes background, provides greater product yields, and requires less optimization than other long and accurate polymerases. *Certamp Complex* was also able to amplify targets with as little as 0.5 pg of template. Additionally, the 10X *Certamp Complex Buffer MgCl₂ FREE* provided with the kit has been optimised for high yields and increased fidelity.

Applications of *Certamp Complex*:

- Amplification of genomic DNA
- Amplification of GC-rich templates
- Amplification of DNA targets with extensive secondary structure
- Amplification from low amounts of templates (≥ 0.5 pg)
- Cloning: PCR product can be used in T/A cloning as well as blunt-ended cloning.

2. STORAGE CONDITIONS

Store all components of the CERTAMP Complex at -20°C, in a **constant temperature freezer** (do not use frost-free freezers). Repeated freeze-thaw cycles do not reduce performance of the product.

3. KIT REAGENTS

Certamp Complex Enzyme Mix (1 U/μl)-Mix of *proof-reading* and non *proof-reading* DNA polymerases.

10X Certamp Complex Buffer MgCl₂ FREE - The buffer has been formulated for optimal PCR fidelity and yield when performing complex PCR. This buffer has adjuncts and stabilizers which allow an easy optimisation of complex amplification reactions. This buffer has been optimised for high yields and increased fidelity in PCR techniques.

10X Reaction Buffer MgCl₂ FREE- 750 mM Tris HCl (pH 9.0), 500 mM KCl, 200 mM (NH₄)₂SO₄. It is an alternative reaction buffer for conventional PCR amplifications.

50 mM MgCl₂ Solution-The reaction buffers included in this product have no Mg²⁺ ion in its composition ; magnesium concentration should be optimised for each amplification reaction.

4. ALTERNATIVE PROTOCOLS

Optimal reaction conditions should be optimal for each experiment. Biotoools recommends two different user protocols for the CERTAMP Complex, according to the individual target system to be amplified.

I. Protocol 1 – 10X Reaction Buffer MgCl₂ FREE

This protocol has been optimised for conventional PCR amplifications.

- 1.- Prepare a Master Mix for the appropriate number of samples to be amplified (see Table 1). The inclusion of positive and negative amplification controls is recommended.
- 2.- Thaw and preserve reagents in ice during their manipulation. Reaction tubes should be stored in ice before thermal cycling.

Table 1: *Master Mix 1*

Reactive	Amount per 25 μl reaction
CERTAMP COMPLEX ENZYME MIX	1 μl
10X REACTION BUFFER MgCl ₂ FREE	2.5 μl
50 mM MgCl ₂ Solution	1 μl
dNTPs 10 mM each (Ref. 20.031)	0.5 μl
Primers	10-20 pmol of each primer
Distilled water	Up to 25 μl

- 3.- Aliquot the Master Mix into PCR tubes and add DNA template.
- 4.- Perform PCR using optimised cycling conditions.
- 5.- Analyse the PCR amplification products on agarose gel.

II. Protocol 2 – 10X Certamp Complex Buffer MgCl₂ FREE

This protocol has been optimised for complex targets.

1.- Prepare a Master Mix for the appropriate number of samples to be amplified (see Table 2). The inclusion of positive and negative amplification controls is recommended.

2.- Thaw reagents and preserve them in ice during their manipulation. Reaction tubes should also be kept in ice before thermal cycling.

Table 2: Master Mix 2

Reactive	Amount per 25 µl reaction
CERTAMP COMPLEX ENZYME MIX	1 µl
10X Certamp Complex Buffer MgCl ₂ FREE	2.5 µl
50 mM MgCl ₂ Solution	1 µl
dNTPs 10 mM each (Ref. 20.031)	0.5 µl
Primers	10-20 pmol of each primer
Distilled water	Up to 25 µl

3.- Aliquot the Master Mix into PCR tubes and add DNA template (0.5 pg to 500 ng).

4.- Perform PCR using optimised cycling conditions.

5.- Analyse the PCR amplification products on agarose gel.

III. Amplification of difficult template (human DNA) using Certamp Complex

CERTAMP Complex gives good results in amplification of human DNA.

1.- Prepare a Master Mix for the appropriate number of samples to be amplified (see Table 3). The inclusion of positive and negative amplification controls is recommended.

2.- Thaw and preserve reagents in ice during their manipulation. Reaction tubes should be store in ice before thermal cycling.

Table 3: Master Mix

Reactive	Amount per 25 µl reaction
CERTAMP COMPLEX ENZYME MIX	1 µl
10X Certamp Complex Buffer MgCl ₂ FREE	2.5 µl
50 mM MgCl ₂ Solution	1 µl
dNTPs 10 mM each (Ref. 20.031)	0.5 µl
Specific primers	0.5 µl (20 µM of each primer)
Distilled water	Up to 25 µl


3.- Aliquot the Master Mix into PCR tubes and add DNA template (0.5 pg to 500 ng).

4.- Perform amplification using the following PCR cycling:

CYCLE STEP	Nº CYCLES	TEMPERATURE	DURATION
Initial Denaturalisation	1	94°C	3 min
Denaturalisation	35	94°C	30 sec
Annealing		55°C	1 min
Extension		72°C	3 min
Final Extension	1	72°C	10 min
Cooling	∞	4°C	∞

5.- Analyse the PCR amplification products on 1-1.5 % agarose gel. The specific product has a size of 3 Kb.

Amplification of different amounts of human DNA using Biotoools DNA Polymerase and Certamp Complex



Biotoools DNA Polymerase

Certamp Complex

Lanes 1, 10: Negative controls (without DNA)

Lanes 2, 11: 100 pg of template DNA

Lanes 3, 12: 15 pg of template DNA

Lanes 4, 13: 7.5 pg of template DNA

Lanes 5, 14: 3.75 pg of template DNA

Lanes 6, 15: 1.875 pg of template DNA

Lanes 7, 16: 0.937 pg of template DNA

Lanes 8, 17: 0.47 pg of template DNA

Lane 9: 1 Kb Ladder marker (Ref. 31.005)

5. TROUBLESHOOTING

Problem	Cause	Solution
No product or low yield	Missing reagent or pipetting error	Check the concentration and storage conditions of dNTPs, primers, etc. Repeat the PCR and make sure that no pipetting errors or missing reagents occur.
	Problems with the template DNA	Check the concentration and quality of the starting material. If the template is difficult e.g. rich in GC sequences we recommend adding DMSO to the master mix. Repeat the PCR with a new dilution of template or with a new DNA purification
	Problems with the primers	Revise the primer design and primer condition. Avoid any designs prone to the formation of primer dimers. Repeat PCR with different primer concentration from 0.1-0.5 µM in 0.1 increments. Check primer degradation on a denaturing polyacrylamide gel.
	Concentration of MgCl ₂	Repeat the PCR with different Mg ²⁺ concentration from 1.5-4 mM in 0.25 increments
	Low enzyme mix concentration	Increase the enzyme- mixture concentration in 0.2 U increments
	PCR cycling non optimal	Check the following parameters of the amplification program: Denaturation - increase the temperature and length of initial denaturation. Annealing - optimise the annealing temperature and time. To increase the specificity perform a touchdown or step-down PCR cycling. Extension time - in case the extension time is short increase the extension time by increments of 30 sec. Number of cycles - i perform additional cycles by increments of 5 cycles. Verify or include a final elongation step.
Nonspecific amplification products or background smear	Annealing temperature too low	Increase the annealing temperature in increments of 1°C.
	Primer design	Review the primer design and primer condition. Both primers must have the same concentration. Titrate concentration of primers. Repeat PCR with different primer concentration from 0.1-0.5 µM in 0.1 increments. Check primer degradation on a denaturing polyacrylamide gel.
	Too much DNA template	Use a serial dilution of the template
	Carryover contamination	Always prepare a negative control PCR (no template). If the NC shows a PCR product or a smear exchange all reagents..
	Enzyme-mix concentration too high	Optimise the concentration of enzyme-mix in your experiment
	PCR cycling non optimal	Use <i>touchdown</i> or <i>stepdown</i> thermocycler. Reduce the number of cycles
PCR products in negative control	Concentration of MgCl ₂	Repeat the PCR with different Mg ²⁺ concentrations from 1.5-4 mM in 0.25 increments
	Carryover contamination	Repeat the PCR exchanging all the reagents.

6. ORDERING INFORMATION

DESCRIPTION	Size	Reference
Certamp Complex Enzyme Mix (1U/µL)	280 µL	10.342
	550 µL	10.343
10X Certamp Complex Buffer MgCl₂ FREE	1.4 mL	10.342
	1.4 mL	10.343
10X Reaction Buffer MgCl₂ FREE	1.8 ml	10.342
	1.8 ml	10.343
50 mM MgCl₂ Solution	1.8 ml	10.342
	1.8 ml	10.343

BIOTOOLS
BIOTOOLS B&M LABS.S.A.